

COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टीट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD

An ISO 17025 : 2017, 9001 : 2015 Establishment  
An Autonomous R&D Organisation under Ministry of Heavy Industries, Govt. of India.  
KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.

☎ 91-491-2569010, 2566120, 2566206 📠 91-491-2566326 ✉ customercare@fcriindia.com 🌐 www.fcriindia.com

Air Flow Laboratory



**CERTIFICATE OF CALIBRATION**

MASS FLOW METER

FOR

M/s. HEKA MEDICALS INDIA PRIVATE LIMITED,  
MONIPPALLY, KOTTAYAM, KERALA.



CC-2395

PAGE 1 OF 5 PAGES

ULR No: CC23952200000386F

DATE OF RECEIPT	DATE OF CALIBRATION	DATE OF ISSUE	RECOMMENDED DUE DATE	CERTIFICATE NUMBER
06.06.2022	06.06.2022	10.06.2022	06.06.2023	CA 373 2206 148

Approved Signatory

सी. बी. सुरेश / C. B. SURESH

प्रमुख अनुसंधान अभियंता / Chief Research Engineer  
फ्लूइड कंट्रोल रिसर्च इंस्टीट्यूट  
Fluid Control Research Institute  
कंजिकोड पश्चिम, पालक्काड-६७८ ६२३  
Kanjikode West, Palakkad-678 623

**SUMMARY**

Device under calibration	: Mass Flow Meter
Model No.	: FS6122-250-F250-5P40-TH1
Serial No.	: B2Q06983
Specified Range	: 0-250 SLPM
Specified Medium	: Air
Make	: Siargo
Laboratory	: Air Flow Laboratory
Reference Standard	: Bell Prover
Calibration Medium	: Compressed Air
Calibration Procedure Code	: WP-AC-12

**RESULTS**

Calibrated Flow Range (SLPM)	Error Range (% FS)	Expanded Uncertainty in Calibration(k=2)
2 to 60	-0.52 to 0.31	Refer Table 2

Calibrated by

Certificate Prepared by

Certificate Checked by

1. Unnikrishnan.P.N

2. Akash M

Akash M

Unnikrishnan.P.N